Supplementary material

Source identification of atmospheric particle-bound metals at Terra Nova Bay, Antarctica

Andrea Bazzano,^A Francesco Soggia^A and Marco Grotti^{A,B}

^ADepartment of Chemistry and Industrial Chemistry, University of Genoa, Via Dodecaneso 31, 16146

Genoa, Italy

^BCorresponding author. Email: grotti@unige.it

Instrument	Varian Inc. (Palo Alto, CA, USA) Vista PRO				
Nahaligan	Valial IIIC. (1 alo Alto, CA, OSA) Vista I KO				
Nebuliser	Pneumatic concentric (K-style)				
Spray chamber	Cyclonic				
Plasma source configuration	Axial				
RF Power	1100 W				
Plasma gas flow rate	15.0 L min ⁻¹				
Auxiliary gas flow rate	1.5 L min ⁻¹				
Nebuliser gas flow rate	0.75 L min ⁻¹				
Sample uptake rate	0.78 mL min ⁻¹				
Internal standard uptake rate	0.22 mL min^{-1}				
Integration time	15 s				
Replicates	7				
Analytes and wavelengths (nm)	Al (236.705; 237.312; 396.152)				
	Fe (234.350; 240.489; 258.588)				
	Lu (291.139) ^A				
	Mg (279.800; 285.213; 293.651)				
	Na (568.821; 588.995; 589.592)				

Table S1. Inductively coupled plasma-atomic emission spectroscopy instrumental parameters

^AInternal standard.

Instrument	Perkin-Elmer (Waltham, MA, USA) Elan DRC II
Nebuliser	PFA-ST microconcentric
Spray chamber	Low-volume cyclonic (cinnabar)
RF Power	1500 W
Plasma gas flow rate	14.5 Lmin^{-1}
Auxiliary gas flow rate	1.65 L min ⁻¹
Nebuliser gas flow rate	0.95 Lmin^{-1}
Sample uptake rate	150 μ L min ⁻¹
RF amplitude	150 V
Axial field voltage	300 V
Cell path voltage	–28 V
Cell rod offset	-8.0 V
Mass analyser rod offset	0.0 V
Stability parameters ^A	RPa = 0.0; RPq = 0.25
Dwell time	50 ms
Sweeps	20
Replicates	5
Measured ions	${}^{7}\text{Li}^{+}, {}^{51}\text{V}^{+}, {}^{52}\text{Cr}^{+}, {}^{55}\text{Mn}^{+}, {}^{59}\text{Co}^{+}, {}^{63}\text{Cu}^{+}, {}^{85}\text{Rb}^{+}, {}^{89}\text{Y}^{+}, {}^{98}\text{Mo}^{+}, {}^{208}\text{Pb}^{+}, {}^{115}\text{In}^{+B}$

 Table S2.
 Inductively coupled plasma-mass spectrometry instrumental parameters

^AMathieu stability parameters of the cells quadrupole: $a = 1.9 \times RPa$; $q = 0.95 \times RPq$;

^BInternal standard.

Table S3. Field blanks and limits of detection														
Stage (µm)	m) Major elements (ng m^{-3})				Trace elements (pg m ⁻³)									
	Al	Fe	Mg	Na	Co	Cu	Li	Mn	Mo	Pb	Rb	Y	Cr	V
Field blanks														
10-7.2	1.26	0.17	0.05	43.3	0.34	2.16	0.08	2.43	0.48	1.50	0.26	0.44	14.0	0.50
7.2-3.0	1.21	0.17	0.05	43.2	0.38	2.60	0.06	3.47	0.41	0.65	0.31	0.47	17.9	0.35
3.0-1.5	1.21	0.17	0.06	43.5	0.36	3.13	0.08	2.77	0.61	0.62	0.23	0.47	16.5	0.43
1.5-0.95	1.21	0.17	0.07	43.5	0.39	2.63	0.10	3.07	0.41	0.73	0.29	0.46	14.3	0.36
0.95-0.49	1.42	0.17	0.09	43.5	0.38	3.01	0.17	2.36	0.40	0.48	0.28	0.49	16.5	0.38
Limits of detection	on													
10-7.2	0.17	0.30	0.22	2.06	0.05	0.20	0.05	1.53	0.10	0.20	0.10	0.03	2.88	0.34
7.2-3.0	0.04	0.30	0.22	2.14	0.01	2.70	0.03	4.64	0.08	0.15	0.09	0.05	6.86	0.04
3.0-1.5	0.08	0.30	0.22	1.96	0.25	6.34	0.04	4.17	0.50	0.26	0.18	0.22	2.15	0.11
1.5-0.95	0.01	0.30	0.21	2.02	0.01	2.23	0.04	4.45	0.03	0.29	0.10	0.01	0.76	0.04
0.95-0.49	0.04	0.29	0.20	2.08	0.05	1.28	0.25	1.72	0.03	0.10	0.05	0.01	1.68	0.15

Table S4. Analysis of CRM-MURST-ISS-A1 (Antarctic sediment)

Values are averages and 95 % confidence intervals, indicative values within parentheses

Element	Unit	Measured	Certified
Al	%	6.45 ± 0.08	6.71 ± 0.33
Со	$\mu g g^{-1}$	6.76 ± 0.76	6.87 ± 0.31
Cr	$\mu g g^{-1}$	46.9 ± 3.5	42.1 ± 3.4
Cu	$\mu g g^{-1}$	5.97 ± 1.01	(5.79 ± 1.59)
Fe	%	2.34 ± 0.13	2.44 ± 0.07
Mg	%	0.96 ± 0.04	(1.52 ± 0.13)
Mn	$\mu g g^{-1}$	437 ± 43	446 ± 19
Мо	$\mu g g^{-1}$	0.9 ± 0.7	(1.1 ± 0.6)
Na	%	2.20 ± 0.16	(2.27 ± 0.06)
Pb	$\mu g g^{-1}$	20.5 ± 1.7	21.0 ± 2.9
Rb	$\mu g g^{-1}$	123.4 ± 9.2	(124.8 ± 5.1)
Y	$\mu g g^{-1}$	18.5 ± 2.4	(19.1 ± 2.5)
V	$\mu g g^{-1}$	51.6 ± 9.9	(51.8 ± 2.2)



Fig. S1. Factor profiles from positive matrix factorisation (PMF) analysis. Values are percentage contribution of each factor.